



EA MLA Signatory  
Český institut pro akreditaci, o.p.s.  
Olšanská 54/3, 130 00 Praha 3

issues

according to section 16 of Act No. 22/1997 Coll., on technical requirements for products, as amended

# CERTIFICATE OF ACCREDITATION

No. 346/2020

**SVCS Process Innovation s.r.o.**  
with registered office **Optátova 708/37, Jundrov, 637 00 Brno, Company Registration**  
**No. 27711170**

to the Calibration Laboratory No. **2393**  
Calibration Laboratory

Scope of accreditation:

Calibration of gas mass flow meters to the extent as specified in the appendix to this Certificate.

This Certificate of Accreditation is a proof of Accreditation issued on the basis of assessment of fulfillment of the accreditation criteria in accordance with

ČSN EN ISO/IEC 17025:2018

In its activities performed within the scope and for the period of validity of this Certificate, the Body is entitled to refer to this Certificate, provided that the accreditation is not suspended and the Body meets the specified accreditation requirements in accordance with the relevant regulations applicable to the activity of an accredited Conformity Assessment Body.

This Certificate of Accreditation replaces, to the full extent, Certificate No.: 184/2017 of 24. 3. 2017, or any administrative acts building upon it.

The Certificate of Accreditation is valid until: **28. 5. 2025**

Prague: 28. 5. 2020



Jiří Růžička  
Director  
Czech Accreditation Institute  
Public Service Company



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**SVCS Process Innovation s.r.o.**  
Calibration Laboratory  
Zámecká 133/78, 757 01 Valašské Meziříčí

**CMC for the field of measured quantity: Flow**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit	max.	unit					
1	Mass flow rate Molbloc-L	1 ml <sub>n</sub> /min	to	10 l <sub>n</sub> /min	Gas	N2	0.2 %	Method of direct comparison with the standard	SVCS KM 1.2002	
		3 l <sub>n</sub> /min	to	30 l <sub>n</sub> /min	Gas	N2	0.3 %			
		10 l <sub>n</sub> /min	to	100 l <sub>n</sub> /min	Gas	N2	0.5 %			
	Molbloc-S	15 l <sub>n</sub> /min	to	1,000 l <sub>n</sub> /min	Gas	N2	0.2 %			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Explanatory notes: SVCS KM internal calibration method

Index "n" at mass flow rate volume units identifies the reference values for temperature T=273.15 °K and pressure p=101325 Pa.

